



Application Guide

Using Sample Bags with the Vac-U-Chamber



The SKC Vac-U-Chamber is a rigid air sample box that allows sample bags to be filled directly by using negative pressure provided by a personal air sampling pump. Both sample and pump contamination are eliminated because the air sample does not pass through the pump. The Vac-U-Chamber features rigid walls that will not collapse under vacuum conditions. All surfaces in contact with the sample are constructed of inert materials.

Required Equipment

1. An **air sampling pump** capable of sampling at the recommended flow rate, such as:
 - SKC Universal Series Sampler
 - SKC AirChek® 52 Sampler
 - SKC AirChek 2000 Sampler
 - SKC AirChek XR5000 Sampler
 - SKC Pocket Pump® Sampler
2. An **air sample bag**, such as:
 - SKC 1-liter Tedlar® Bag Cat. No. 232-01 for use with small Vac-U-Chamber
 - SKC 1-liter SamplePro® FlexFilm Bag Cat. No. 236-001 for use with small Vac-U-Chamber
 - SKC 8-liter FlexFilm Bag Cat. No. 236-004 for use with large Vac-U-Chamber
3. **SKC Vac-U-Chamber:**
 - Small Cat. No. 231-940
 - Large Cat. No. 231-939

Optional Equipment

1. **Nitrogen** (99.999% pure)
2. An **airflow calibrator**, such as:
 - Defender Primary Standard Calibrator 717 Series

Note: Calibration is required when sampling according to a method.

Introduction

Figure 1 shows the SKC large Vac-U-Chamber (Cat. No. 231-939) with an air sample bag in a sampling train with a Universal Series Sampler.

To determine the correct flow rate for the chemical of interest, refer to the appropriate analytical method. Consult the sample pump operating instructions for further details on using the pump.




Figure 1.

Universal Series Sampler connected to a large Vac-U-Chamber for sampling

1. Calibrating Pump Flow Rate

The pump must be calibrated if taking a bag sample according to an analytical method that specifies a flow rate. **Note:** Calibration is not necessary if simply taking a grab sample. Care should be taken to ensure the bag is not filled more than 80% of its maximum volume (Figure 2).




Figure 2. Bag inflation

Ensure pump has run for 5 minutes before calibrating. Use flexible tubing to connect the pump inlet to the outlet (suction) port of a calibrator. **Note:** If using a Universal Series Sampler, ensure that the pump is in the high flow mode. See pump operating instructions.

Using Sample Bags with the Vac-U-Chamber

Calibrate the pump to the flow rate specified in the analytical method. Refer to pump and calibrator operating instructions for details on calibrating pump flow rate or request SKC Application Guide: *Calibrating a Pump Using a Film Flowmeter* (SKC Publication #1163) or *Calibrating a Pump Using an Electronic Calibrator* (SKC Publication #1366).

2. Assembling the Sampling Train

Connect the inlet stem of an appropriate size bag to the sample inlet (PTFE tubing) inside the Vac-U-Chamber (Figure 3) or use the PTFE tube adapter for bags with a vertical valve (231 and 237 Series bags). Open the valve on the bag. Refer to the bag operating instructions.




Figure 3.
Ports inside a small
Vac-U-Chamber

Purge Port (Purge port located on front of large Vac-U-Chamber)

3. Purging the Sample Bag

Attach the quick-connect fitting on the end of the supplied Tygon® tubing to the purge port of the Vac-U-Chamber. Attach the open end of the tubing to the pump inlet (Figure 4). Before evacuating the sample bag, ensure the red cap plug remains on the sample inlet port outside the Vac-U-Chamber. Activate the pump and allow it to run until the bag is completely evacuated. Turn off the pump.

4. Purging the Sample Line

To purge the sample line, remove the red cap plug and connect a long length of PTFE tubing to the sample inlet port on the outside of the Vac-U-Chamber (Figure 4). Connect the loose end of the tubing to a clean air source. Turn on the pump for a period adequate to purge the line. Repeat as needed.



Figure 4. Universal
Series Sampler
connected to a small
Vac-U-Chamber for
purging a bag or
purging the sample
line

Red Cap Plug
(Leave on for bag evacuation,
remove for line purge)

Sample Inlet Port
(Connect purified air source here for
line purge.)

5. Filling the Bag with a Sample

With the pump still attached to the Tygon tubing, detach the quick-connect fitting from the purge port on the Vac-U-Chamber and insert it into the vacuum port (Figures 1 and 5). Close the Vac-U-Chamber and secure both latches. Ensure the decompression valve located below the handle is closed (Figure 5).




Figure 5.
Ports outside
a small
Vac-U-Chamber

Grab Sampling

Activate the sampling pump and view the bag inflating through the window on top (Figure 1). Allow the pump to run until the bag is approximately 80% full. **Do not over-inflate the bag (Figure 2).** Turn off the pump and immediately open the Vac-U-Chamber. Close the valve on the bag. If the Vac-U-Chamber is difficult to open after sampling, loosen the decompression valve to relieve the pressure produced during sampling.

Timed Sampling

Activate the pump calibrated to the recommended flow rate and record the start time. Sample for the time specified in the analytical method. Turn off the pump, record the stop time, and close the valve on the bag.

Samples can be analyzed directly from the bag using a color detector tube, or the sample can be sent to a laboratory for analysis.

6. Shipping Bag Samples

Sample bags sent to a laboratory for analysis should be packed loosely and padded to minimize possible puncturing during shipment. Bag samples should not be shipped by air unless the cargo cabin is pressurized. A significant decrease in barometric pressure may cause sample bags to burst.

Copyright 2001 - 2011

Notice: This publication is intended for general information only and should not be used as a substitute for reviewing applicable government regulations, equipment operating instructions, or legal standards. The information contained in this document should not be construed as legal advice or opinion nor as a final authority on legal or regulatory procedures.