



**FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND**  
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

## ***Declaration of Conformity***

**Extech Model: EA33**

**Description: EasyView Light Meter with Memory**

**Date of Issue: 06-Sep-13**

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

**EMC Directive: 2004/108/EC**

**Report Number: 1303015E-01**

**Report Date of Issue: 8/1/2013**

**Standards:**

CISPR 11:2009+A1:2010 & EN 55011:2009+A1:2010

EN 61326-1:2006

IEC 61000-4-2:2008

IEC 61000-4-3:2006+A1:2007+A2:2010

AS/NZS CISPR 11:2011

The test reports show that the product fulfills the requirement in the EC EMC Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive.

A handwritten signature in black ink, appearing to read "Mark Sultzbach", written over a horizontal line.

Mark Sultzbach (QA Administrator)  
(for Tony Campagna Director of Quality Services)